



ULTRA FOR MEDICAL PRODUCTS CO.

Declaration of Conformity **Silicon Foley Catheter**

- Manufacturer's Name** : Ultra For medical products Co (U.M.I.C).
Manufacturer's Address : Part No. (304:310) & part no. (312) – Industrial Area, Arab El Awamer – Abnoub - Assiut – Egypt.
Intended use : The Foley Catheter is a tubular device intended for being introduced into the urinary bladder through the urethra in order to provide drainage, drug delivery and/or flushing of the bladder.
Internet : www.ultramedumic.com
Classification Rule : Class IIa - Rule no. 5 (Annex-IX) in accordance of the Council Directive MDD 93/42/EEC of 14 June 1993 amended by Directive 2007/47/EC, September 2007.
Authorized Representative (AR)
Name : Obelis^{SA}
Address : Bd. Général Wahis 53
1030 Brussels - Belgium
Phone : +32 (0)2 732 59 54
Fax : +32 (0)2 732 60 03
E-mail : mail@obelis.net
Internet : www.obelis.com

We hereby declare that the distributed CE marked products, specified in the attached product list, conform to the type covered by the provisions of the Council Directive 93/42/EEC, June 1993 concerning medical devices amended by Directive 2007/47/EC, September 2007 under the supervision of the Notified Body "HT CERT" (CE 2803)".

- Based on**
Certificate No. : 1421C05201201
Issue Date : 14/12/2020
Expiry Date : 26/05/2024

G.F.I. Health Technology Certification Ltd

- Address** : Jakovides Tower 81 – 83 Grivas Digenis Av., CY 1090, Nicosia Cyprus
Phone : +357 (22) 503000
Fax : +357 (22) 503001
Internet : www.htcert.com

Declaration of Conformity Sterile Silicon Foley Catheter _TF_19_01

ISSUE/REV: - 4/0

Page 1 of 2

**ULTRA
MED****شركة التيرا
للمنتجات الطبية****ULTRA FOR MEDICAL PRODUCTS CO.**

In addition, we ensure and declare that the distributed CE marked products, as mentioned and falling within **Class IIa** , meet the provisions of the EC-Directive that apply to them.

This declaration is based on MDD 93/42/EEC Annex II (- Sect. 4) of directive 93/42/EEC, June 1993 amended by Directive 2007/47/EC, September 2007.

Attached List:

Code	Item
3007-6/2	Silicon Foley Balloon Catheter Size 6 FG / 2-Way
3007-8/2	Silicon Foley Balloon Catheter Size 8 FG / 2-Way
3007-10/2	Silicon Foley Balloon Catheter Size 10 FG / 2-Way
3007-12/2	Silicon Foley Balloon Catheter Size 12 FG / 2-Way
3007-14/2	Silicon Foley Balloon Catheter Size 14 FG / 2-Way
3007-16/2	Silicon Foley Balloon Catheter Size 16 FG / 2-Way
3007-18/2	Silicon Foley Balloon Catheter Size 18 FG / 2-Way
3007-20/2	Silicon Foley Balloon Catheter Size 20 FG / 2-Way
3007-22/2	Silicon Foley Balloon Catheter Size 22 FG / 2-Way
3007-24/2	Silicon Foley Balloon Catheter Size 24 FG / 2-Way
3007-14/3	Silicon Foley Balloon Catheter Size 14 FG / 3-Way
3007-16/3	Silicon Foley Balloon Catheter Size 16 FG / 3-Way
3007-18/3	Silicon Foley Balloon Catheter Size 18 FG / 3-Way
-20/3	Silicon Foley Balloon Catheter Size 20 FG / 3-Way
3007-22/3	Silicon Foley Balloon Catheter Size 22 FG / 3-Way
3007-24/3	Silicon Foley Balloon Catheter Size 24 FG / 3-Way


General Manager

Dated on: - 20/12/2012



Assiut Factory: Part No. (304:310) & part no. (312) – Industrial Area, Arab El Awamer – Abnoub - Assiut – Egypt. **Tel:** 002-088-4964333 (500), 002-088-4964666 (600), 002-088-4964777 (700), 002-088-4964888 (800), 002-088-4964999 (900)

& Fax:002-088-4964222**& Mob:** 002-01001558853, 002-01068832355

Cairo Head Office: 64, Nakhla El Motaiy Triumph Square Heliopolis **Tel:** 022/4171621-4143794 **& Fax:** 022/4171613 **& Mob:** 01223988200

Assiut Office: 23, July Str. **Tel:** 088/2364111 – 2364222 **& Fax:** 088/2334964 **& Mob. :** 01223988202

Alexandria Office: 212Abd El salaam Airef Str, Luran **Tel:** 03/5856202– 5856458 **& Fax:** 03/5828988**& Mob:** 01223948666

E-mail: ultra@elaggargroup.com, shady@elaggargroup.com

Website: www.ultramedumic.com

Declaration of Conformity Sterile Silicon Foley Catheter _TF_19_01

ISSUE/REV: - 4/0

Page 2 of 2